

Xylem Analytics Germany GmbH
Am Achalaich 11
D-82362 Weilheim
Germany



UK - DECLARATION OF CONFORMITY

| |
|--|
| We declare under our sole responsibility that the product described herein |
| TITRONIC 500 |
| is in conformity with the requirements of the following documents: |

| |
|---|
| Electromagnetic Compatibility Regulation 2016 BS EN 61326-1:2013 |
| Electrical Equipment (Safety) Regulation 2016 BS EN 61010-1:2010 |
| The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulation 2012 BS EN IEC 63000: 2018 |
| Radio Equipment Regulation 2017 |

Mainz, 22.08.2022
Conf. No.: UK-Titrat011


Robert Reining
Managing Director



SI Analytics



are brands of Xylem Inc. or one of its subsidiaries.